

ENTRY NO. 116

NAME OF MACHINE Davis 76" Cyclotron
 INSTITUTION Crocker Nuclear Laboratory
 ADDRESS Davis, California, USA
 TEL (916) 762-1460 TELEX
 IN CHARGE Thomas A. Cahill REPORTED BY CNL Personnel

HISTORY AND STATUS

DESIGN, date Model tests
 ENG DESIGN, date ORIC Copy
 CONSTRUCTION, date 1964-1966
 FIRST BEAM, date (or goal) 1966
 MAJOR ALTERATIONS none

COST, ACCELERATOR 1.4×10^6
 COST, FACILITY, total 4.5×10^6

FUNDED BY . Recharges for beams and services

ACCELERATOR STAFF, OPERATION AND DEVELOPMENT

SCIENTISTS 1.5 ENGINEERS 1.0

TECHNICIANS 3.0 CRAFTS 2.0

GRAD STUDENTS involved during year 3.0

OPERATED BY Research staff or 3.0 Operators

OPERATION 140 hr/wk. On target 50 hr/wk

TIME DISTR. in house 60 %, outside 40 %

BUDGET, op & dev 500. K\$

FUNDED BY . Recharges for beams and shops

RESEARCH STAFF, not included above

USERS, in house 12 outside 24

GRAD STUDENTS involved during year 40

RESEARCH BUDGET, in house 1.0 M\$

FUNDED BY . NPS, EPA, DOE, NSF

MAGNET

POLE FACE, diameter (compact) 193 cm, R-extraction 80 cm

R injection 19 cm

GAP, min 19 cm, Field 22.7 kG

max 71 cm, Field 12.7 kG at 0.8×10^6

AVERAGE FIELD at R ext 17.5 kG Ampere turns

B max/ $\langle B \rangle$ 13

NUMBER OF SECTORS { compact 3 } Spiral, max 30 deg

SECTOR ANGLE (SSC) deg

TRIMMING COILS 10

CONDUCTOR, material and type Hollow copper

STORED ENERGY (cryogenic) MJ

POWER: main coils 800 max kW: current stability 5×10^{-4}

trimming coils 350 max kW: current stability 5×10^{-4}

WEIGHT: Fe 268 tons: coils 42 tons

COOLING system deionized water

ION ENERGY (Bending limit) E/A = 90 q^2/A^2 MeV/amu

(Focusing limit) E/A = q/A MeV/amu

ACCELERATION SYSTEM

DEES, number 1 angle 180 deg

BEAM APERTURE 4.5 cm; DC Bias 0 kV

TUNED by, coarse coarse short plane trim capacitors

RF 7.3 to 22.5 MHz, stable $\pm 1/10^6$

Orb F 1.5 to 22.5 MHz

HARMONICS, RF/Orb F, used 1.3

DEE-Gnd, max 60 kV, min gap 215 cm

STABILITY, (pk-pk noise)/(pk RF volt) 0.005

ENERGY GAIN, max 120 kV/turn

RF PHASE, stable to ± 10 deg

RF POWER input, max 150 kW

FREQUENCY MODULATION, rate /s

modulator, type

beam pulse, width

VACUUM SYSTEM

OPERATING PRESSURE 10^{-6} Torr or mbar

PUMPS, No, Type, Size 2 diffusion, 2m

ION SOURCES

Hot filament, modified LBL 88" source

INJECTION SYSTEM

none

EXTRACTION SYSTEM

Electrostatic + 2 magnetic

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 360 m²; movable m²

TARGET STATIONS 10 in rooms

STATIONS served at same time, max 1

MAG SPECTROGRAPH type none

COMPUTER model PDP 15/40; 2 PDP 11/44; PDP 11/23

OTHER FACILITIES XRF, PIXE systems; Co-60 irradiation source; hot radiochemical labs; isotope production; large area irrads. station; O-(n,p) facility; off-line counting station (ND65 &

CHARACTERISTIC BEAMS ND66 multichannel analyzers)

PARTICLE	ENERGY (MeV)	CURRENT (p μ A)
Goal	Achieved	Internal

P	4 to 68	30.0
---	---------	------

d	15 to 45	40.0
---	----------	------

α	16 to 90	40.0
----------	----------	------

^{3}He	20 to 90	1.0
-----------------	----------	-----

SECONDARY	15 to 65	10 ⁶ (part/s)
-----------	----------	--------------------------

n	15 to 65	10 ⁶
---	----------	-----------------

BEAM PROPERTIES

MEASURED	CONDITIONS
----------	------------

PULSE WIDTH 8 RF deg	20 p μ A of 67.5 MeV p ions
----------------------------	---------------------------------------

PHASE EXC, max 42 RF deg	10 p μ A of 50.0 MeV ions
--------------------------------	-------------------------------------

EXTRACT eff 90 %	20 p μ A of 67.5 MeV p ions
------------------	---------------------------------------

RESOL ΔE/E 0.4 %	20 p μ A of 67.5 MeV p ions
------------------	---------------------------------------

EMITTANCE	axial
-----------	-------

(π mm-mrad)	p rad
------------------	-------

	p μ A of MeV
--	------------------------

OPERATING PROGRAMS	time distribution
--------------------	-------------------

BASIC NUCLEAR PHYSICS 10%	SOLID STATES PHYSICS 30%
---------------------------	--------------------------

BIOMEDICAL APPLICAT. 5%	ISOTOPE PRODUCTIONS 15%
-------------------------	-------------------------

Atomic 5%, Analytical Sv. (PIXE) 30%	Biological (N-13) 5%
--------------------------------------	----------------------

REFERENCES/NOTES

- 1) Accelerator supported on beam and services recharge since 1971.

PLAN VIEW OF FACILITY, COMMENTS, ETC.